

Current Status of Light Source and Beamlines

Light Source in 2006

Masahiro KATOH

UVSOR Facility, Institute for Molecular Science

1. Status of UVSOR-II

In the fiscal year 2006, we have operated the UVSOR-II accelerators from April '06 to March '07. We had one 9 week shut-down period in July and August '06 and two 2 week ones in October and March '07. The long shut down in summer was to replace the magnet power supply of the booster synchrotron and to install a new beam duct at a straight section between B6 and B7 for new undulator. Construction of new radiation shield wall was also done during this period. The short shut down in October was to install a new variably polarized undulator. The shut down in March is to replace the bending magnet power supply of the beam transport line, preparing for full energy injection. We also stopped the operation for one week in the holiday week of May and for two weeks around the New Years day.

Fortunately, we had very few troubles on the accelerators during FY2006. We only had a few minor troubles on a kicker magnet power supply of the storage ring and the RF amplifier of the linear accelerator. All of them did not affect the operation schedule.




Figure 1 Monthly statistics of the operation time (left) and of the integrated beam current (right).

We had totally 31 weeks for the users operation, 29 weeks in multi-bunch mode and two weeks in single bunch mode. We had five weeks dedicated for machine studies. The monthly statistics of the operation time and the integrated beam current are shown in Figures 1. The normal operation pattern in a week is as follows. From Tuesday to Friday, the machine is operated for users. The beam injection is twice a day, at 9:00 and 15:00. The initial beam current of each run is 350 mA in multi-bunch mode and 100 mA in single bunch mode. On Monday, the machine is operated for machine studies.

2. Improvements

Commissioning of New Magnet Power Supplies for Booster Synchrotron

The magnet power supplies of the booster synchrotron had been working for more than twenty years. In these years, we had a few very serious troubles on then. In July, 2006, these power supplies were finally replaced. New power supplies have capability of accelerating the electron beam up to the full energy of the storage ring, 750 MeV. In August, we have commissioned the power supplies with the 600 MeV acceleration mode. In March, the power supply for the bending magnets of the beam transport line has been replaced. New power supply is compatible with the full energy injection. In April, 2007, the full energy acceleration, transportation and injection will be started.



Figure 2 New Magnet Power Supplies for Booster Synchrotron

Progress in Orbit Stabilization

The electron orbit of UVSOR-II shows drift motion of a few hundred microns with a time scale of hours. To suppress the orbit drift, a feedback system is under development. As the first step, a sub-system has been constructed to correct the orbit displacement in the horizontal plane automatically by controlling the RF frequency [1]. This system has been successfully commissioned and the orbit drift in the horizontal plane is reduced significantly. The system has been working for about one year and the change of the RF frequency has been recorded, which is strongly correlated with the temperature of the storage ring room floor [1].

As the second step, a sub-system to monitor the positions of the beam position monitors has been developed and tested [2]. The synchrotron irradiation cause thermal deformation of the beam ducts, which results in displacements of the beam position monitors mounted on them. The measurement shows that most of the beam position monitors show the displacements of a few tens of microns. So at present, this is not a serious problem but, in future, when we try to stabilize the orbit within ten microns, it should be considered.

New Undulator for BL7U

A new undulator was installed in the straight section between B6 and B7 in October. This undulator will provide VUV light of linear polarization in both horizontal and vertical planes. It can also provide circular polarized

VUV light. The configuration of the magnet array is of APPLE-II type [3]. The main parameters are shown in Table 1. The undulator is now under commissioning and partly opened for users.

Before installing the new undulator, another undulator of in-vacuum type, which had occupied a part of the straight section, was moved to another short straight section between B5 and B6. It will provide soft X-rays to BL6U, which is now under designing.



Figure 3 New APPLE-II undulator for BL7U (upper) and In-vacuum undulator for BL6U (lower)

3. Researches and Developments

Free Electron Laser

In FY2006, one watt lasing in deep UV region (230nm) was successfully demonstrated [4]. The lower emittance and the higher peak current of the electron beam in UVSOR-II made this possible. Several user experiments have been started. One example is related to the magnet circular dichroism which is described in these reports [5].

Coherent harmonic generation (CHG) is a method to produce coherent radiation in shorter wavelength region where no good mirror is available for the optical cavity. By using the Ti:Sa laser introduced for the bunch slicing experiment, the CHG was successfully demonstrated in collaboration with French group [6]. Coherent third harmonics of the injected laser was clearly observed.

Laser Bunch Slicing

By using a TiSa laser of 1 kHz repetition rate and 2.5 mJ pulse energy, we have succeeded in producing intense coherent terahertz radiation [7]. The short pulse laser interacts only with a part of the electron bunch inside an undulator. There is produced strong energy modulation at the interacting part. As the electron bunch is proceeding in the storage ring, the modulated part is separated from the bunch and there is created a dip on the

bunch. This fine structure on the longitudinal density distribution is the origin of the intense coherent radiation.

[1] K. Suzumura et al., in this report

[2] K. Suzumura et al., in this report

[3] S. Sasaki, Nucl. Instr. Meth. A347 (1994) 83


[4] M. Hosaka et al., in this report

[5] T. Nakagawa, T. Yokoyama, M. Hosaka, M. Katoh, in this report

[6] M. Labat et al., in this report

[7] M. Shimada et al., in this report

UVSOR Accelerator Complex



Parameters of Injection Linear Accelerator

Energy	15 MeV
Length	2.5 m
Frequency	2856 MHz
Accelerating RF Field	$2\pi/3$ Traveling Wave
Klystron Power	1.8 MW
Energy Spread	~ 1.6 MeV
Repetition Rate	2.6 Hz


Parameters of Booster Synchrotron

Energy	600 / 750 MeV
Injection Energy	15 MeV
Beam Current	32 mA (uniform filling)
Circumference	26.6 m
RF Frequency	90.1 MHz
Harmonic Number	8
Bending Radius	1.8 m
Lattice	FODO \times 8
Betatron Tune	
Horizontal	2.25
Vertical	1.25
Momentum Compaction	0.138
Repetition Rate	2.6 / 1 Hz (600 / 750 MeV)

Parameters of UVSOR-II Storage Ring


Energy	750 MeV
Injection Energy	600 / 750 MeV
Maximum Stored Current	500 mA (multi bunch) 100 mA (single bunch)
Natural Emittance	27.4 nm-rad
Circumference	53.2 m
RF Frequency	90.1 MHz
Harmonic Number	16
Bending Radius	2.2 m
Lattice	Extended DBA \times 4
Straight Section	(4 m \times 4) + (1.5 m \times 4)
RF Voltage	100 kV
Betatron Tune	
Horizontal	3.75
Vertical	3.20
Momentum Compaction	0.028
Natural Chromaticity	
Horizontal	-8.1
Vertical	-7.3
Energy Spread	4.2×10^{-4}
Natural Bunch Length	108 ps

Electron Beam Optics of UVSOR-II Storage Ring



Horizontal/vertical betatron functions and dispersion function (left), and horizontal/vertical electron beam sizes and beam divergences (right) of UVSOR-II

Parameters of the insertion devices



Brilliance of radiation from the insertion devices (U3, U5, U6 and U7) and a bending magnet of UVSOR-II

BL3U In-vacuum Undulator

Number of Periods	50
Period Length	38 mm
Pole Length	1.9 m
Pole Gap	15~40mm
Deflection Parameter	2.00~0.24

BL6U In-vacuum Undulator

Number of Periods	26
Period Length	36 mm
Pole Length	0.94 m
Pole Gap	15~40 mm
Deflection Parameter	1.78~0.19

BL5U Helical Undulator / Optical Klystron

Number of Periods	21 / 9+9(Opt. Kly.)
Period Length	110mm
Pole Length	2.35 m
Pole Gap	30~150 mm
Deflection Parameter	4.6~0.07(Helical) 8.5~0.15(Linear)

BL7U Apple-II variable polarization undulator

Number of Periods	40
Period Length	76mm
Pole Length	3.04 m
Pole Gap	24~200 mm
Deflection Parameter	5.4 (max. horizontal) 3.6 (max. vertical) 3.0 (max. helical)

Bending Magnets

Bending Radius	2.2 m
Critical Energy	425 eV

Beamlines in 2006

Eiji SHIGEMASA

UVSOR Facility, Institute for Molecular Science

Eight bending magnets and three insertion devices are available for utilizing synchrotron radiation at UVSOR. There is a total of thirteen operational beamlines in 2006, which are classified into two categories. Eight of them are so-called "Open beamlines", which are open to scientists of universities and research institutes belonging to the government, public organizations, private enterprises and those of foreign countries. The rest of the five beamlines are so-called "In-house beamlines", and are dedicated to the use of research groups within IMS. We have one soft X-rays station equipped with a double-crystal monochromator, seven extreme ultraviolet and soft X-rays stations with a grazing incidence monochromator, three vacuum ultraviolet stations with a normal incidence monochromator, one infrared (IR) station equipped with Fourier-Transform interferometers, one station with a multi-layer monochromator, and three non-monochromatized stations for irradiation of white-light, as shown in the appended table for all available beamlines at UVSOR.

The upgrade project of the UVSOR storage ring, in which the creation of four new straight sections and the achievement of much smaller emittance (27 nm-rad) were planned, has been approved in the fiscal year of 2002 and has been accomplished on schedule. Keeping pace with the upgrade project, the improvements and upgrades of the beamlines at UVSOR have been continuously discussed with users in a series of UVSOR workshops. From a viewpoint of the radiation safety, the experiments carried out at the experimental stations on the second floor such as BL3B and BL7B, and the fine tunings of the laser system installed inside the shield walls during the machine study, seem to become quite risky due to the introduction in the near future of the so-called top-up operation of the UVSOR storage ring. Accordingly we have decided to put two old beamlines, BL8A and BL3B, out of service until the middle of March 2006. In order to transfer the experimental endstation for BL7B to the space after the removal of BL8A, the optical path downstream the exit slit has been changed and a new refocusing mirror system has been installed. It was found through the alignment work at BL7B that the first mirror with a torodial shape had been misaligned seriously, which essentially degrades the performance of the monochromator. It may take several weeks to accomplish the alignment of this mirror, since the change of the position of the first mirror leads to the realignment of all the optical elements of the beamline. We have planned to make the realignment of BL7B in April and May of 2007. The laser system has been moved to the empty lot of BL3B by the machine group.

Regarding the utilization for the long straight section between B06 and B07, a UVSOR workshop has been held in March 2005. A high resolution and high flux variable polarization beamline for spectroscopy in the VUV range has been proposed and possible scientific cases performed on this beamline have been discussed there. The construction of the new beamline BL7U has been completed during the regular summer shutdown in 2006 as had planned. BL7U is composed of a modified Wadsworth-type monochromator with three interchangeable spherical gratings ($R=10$ m; 1200, 2400, and 3600 lines/mm), and a hemispherical photoelectron analyzer, where high-resolution angle-resolved photoemission experiments can be performed. A new APPLE-II type undulator for the light source of BL7U has successfully been installed at the end of October 2006. It has been confirmed that the total performance of BL7U, after careful tunings of the monochromator and the analyzer, nearly reaches the theoretically expected level. Concerning the utilization of the first in-vacuum type undulator, which has been relocated from the long straight section U7 to the short one between B05 and B06, a new project for constructing the undulator beamline BL6U is just beginning execution. A preliminary design study for the

monochromator at BL6U has been terminated, thanks to the collaboration with KEK-PF. The monochromator will cover the photon energy range from 30 to 500 eV, with the resolving power higher than 10000 and the photon flux more than 10^{10} photons/sec. The construction of BL6U is expected to start from the spring of 2008. Further serious discussion toward utilizing the available straight sections most effectively and formulating a basic plan on the beamline construction, will be continued.

All users are required to refer to the beamline manuals and the UVSOR guidebook (latest revision in 1999), on the occasion of conducting actual experimental procedures. Those wishing to use the open and in-house beamlines are recommended to contact the stationmaster/supervisor and the representative, respectively. For updated information of UVSOR, <http://www.uvsor.ims.ac.jp/>.

Station Masters and Supervisors of Open Beamlines in FY2006


Beamline	Station Master	Sub Master	Supervisor
1A	N. Kondo	E. Shigemasa	E. Shigemasa
1B	M. Hasumoto	S. Kimura	S. Kimura
5U	T. Ito	S. Kimura	S. Kimura
5B	M. Hasumoto	E. Nakamura	E. Shigemasa
6B	S. Kimura	E. Nakamura	S. Kimura
7B	M. Hasumoto	S. Kimura	S. Kimura
8B1	Y. Hikosaka	E. Nakamura	E. Shigemasa
8B2	R. Sumii	E. Nakamura	S. Kimura

Station Masters and Supervisors of In-House Beamlines in FY2006

Beamline	Representative	Affiliation
2B	K. Mitsuke	Dep. VUV Photoscience
3U	N. Kosugi	Dep. VUV Photoscience
4A1/4A2	T. Urisu	Dep. VUV Photoscience
4B	E. Shigemasa	UVSOR

Beamlines at UVSOR-II

Beam-line	Monochromator, Spectrometer	Energy Region (eV)	Experiments	Beamline master
1A	Double-Crystal	600 eV – 4 keV	Solid (Absorption)	N. Kondo nkondo@ims.ac.jp
1B	1m Seya-Namioka	1.9 eV – 40 eV	Solid (Reflection, Absorption)	M. Hasumoto hasumoto@ims.ac.jp
2B	18m Spherical Grating (Dragon)	20 eV – 200 eV	Gas (Photoionization, Photodissociation)	K. Mitsuke mitsuke@ims.ac.jp
3U	Varied-Line-Spacing Plane Grating (Monk-Gillieson)	40 eV – 600 eV	Gas (Photoionization, Photodissociation) Solid (Photoemission)	T. Hatsui hatsui@ims.ac.jp
4A1	Multi-Layered-Mirror	50 eV – 95 eV	Irradiation	T. Urisu urisu@ims.ac.jp
4A2	None		Irradiation	T. Urisu urisu@ims.ac.jp
4B	Varied-Line-Spacing Plane Grating (Monk-Gillieson)	25 eV – 800 eV	Gas (Photoionization, Photodissociation) Solid (Photoemission)	E. Shigemasa sigemasa@ims.ac.jp
5U (FEL)	None (Optical Klystron)		Free Electron Laser	J. Yamazaki yamazaki@ims.ac.jp
5U	Spherical Grating (SGM-TRAIN*)	5 eV – 250 eV	Solid (Photoemission)	T. Ito tito@ims.ac.jp
5B	Plane Grating	5 eV – 600 eV	Calibration Solid (Absorption)	M. Hasumoto hasumoto@ims.ac.jp
6B (IR)	Martin-Puplett FT-FIR Michelson FT-IR	0.25 meV – 2.5 eV	Solid (Reflection, Absorption)	S. Kimura kimura@ims.ac.jp
7U	10m Normal Incidence (Modified Wadsworth)	6 eV – 40 eV	Solid (Photoemission)	S. Kimura kimura@ims.ac.jp
7B	3m Normal Incidence	1.2 eV – 30 eV	Solid (Reflection, Absorption)	M. Hasumoto hasumoto@ims.ac.jp
8B1	15m Constant Deviation Grazing Incidence	30 eV – 600 eV	Solid (Absorption)	Y. Hikosaka hikosaka@ims.ac.jp
8B2	Plane Grating	1.9 eV – 150 eV	Solid (Photoemission)	R. Sumii sumii@ims.ac.jp



* Spherical Grating Monochromator with Translating and Rotating Assembly Including Normal incidence mount

BL1A

Soft X-Ray Beamline for Photoabsorption Spectroscopy

BL1A is a soft X-ray beamline for photoabsorption spectroscopy. The beamline is equipped with a focusing premirror and a double crystal monochromator [1]. The monochromator serves soft X-rays in the energy region from 585 to 4000 eV by using several kinds of single crystals such as β -Al₂O₃, beryl, KTP (KTiOPO₄), quartz, InSb, and Ge. The throughput spectra measured by a Si photodiode (AXUV-100, IRD Inc.) are shown in Fig. 1. Typical energy resolution ($E/\Delta E$) of the monochromator is about 1500 for beryl and InSb. There are no experimental setups specific of this beamline, except for a small vacuum chamber equipped with an electron multiplier (EM) detector. Photoabsorption spectra for powdery samples are usually measured in a total electron yield mode, with the use of the EM detector.

[1] Hiraya et al, Rev. Sci. Instrum. **63** (1992) 1264.




Fig. 1 Throughput spectra of the double crystal monochromator at BL1A.



Fig. 2 A side view of BL1A.

Beamline Specifications

Monochromator	Double crystal monochromator
Monochromator crystals: (2 θ value, energy range)	β -Al ₂ O ₃ (22.53 Å, 585-1609 eV), beryl (15.965 Å, 826-2271 eV), KTP (10.95 Å, 1205-3310 eV), quartz (8.512 Å, 1550-4000 eV), InSb (7.481 Å, 1764-4000 eV), Ge (6.532 Å, 2094-4000 eV)
Resolution	$E/\Delta E = 1500$ for beryl and InSb
Experiments	Photoabsorption spectroscopy

BL1B

Seya-Namioka Monochromator for General Purposes

BL1B has been constructed to perform various spectroscopic investigations such as absorption, reflectivity, and luminescence in condensed matters. This beamline consists of a pre-focusing mirror, a 1-m Seya-Namioka type monochromator, and post-focusing mirrors with different focal lengths. Three gratings of 600, 1200, and 2400 l/mm can cover the wavelength region ranging from 40 to 650 nm ($h\nu = 2 - 30$ eV). The post mirror with a longer focal length is usually used with an LiF window to separate the vacuum condition of the monochromator from a main experimental station, which make experiments for liquids and bio-specimens possible, while the other is mainly utilized for solid-state spectroscopy. The output flux from this monochromator is about 10^{10} photons/sec. around 200 nm with 0.1 mm slit openings. The spectral distributions for two gratings measured by a conventional photomultiplier are shown in Fig. 1. A second monochromator (Spex 270M) and a LN-cooled CCD detector (Princeton Inc.) are available for luminescence measurements, together with a liquid helium-flow type cryostat. To perform time-resolved experiments, a TAC system is also available.




Fig. 1 Photocurrent at the sample position at BL1B.




Fig. 2 Photo of BL1B.

Beamline Specifications

Monochromator	1-m Seya-Namioka type
Wavelength Range	40 to 600 nm (2-30 eV)
Resolution	$E/\Delta E \sim 1000$ at 100 nm (10 eV)
Experiment	Absorption, reflection, luminescence spectroscopy for solids

BL2B

Beamline for Gas Phase Photoionization and Reaction Dynamics

This beamline has been developed for the purpose of studying ionization, excitation and decay dynamics involving inner-valence electrons, $2p$ electrons of the third row atoms, and $4d$ electrons of the lanthanides. The monochromator is a spherical grating Dragon-type with 18-m focal length. High throughput (1×10^{10} photons s^{-1}) and high resolution ($E/\Delta E = 2000 - 8000$) are achieved simultaneously under the condition of the ring current of 100 mA [1]. The optical system consists of two prefocusing mirrors, an entrance slit, three spherical gratings (G1 - G3), two folding mirrors, a movable exit slit and a refocusing mirror [2]. The monochromator is designed to cover the energy range of 23 - 205 eV with the three gratings: G1 (2400 lines mm^{-1} , $R = 18$ m) at 80 - 205 eV; G2 (1200 lines mm^{-1} , $R = 18$ m) at 40 - 100 eV; G3 (2400 lines mm^{-1} , $R = 9.25$ m) at 23 - 50 eV. The percentage of the second-order light contamination at $h\nu = 45.6$ eV is 23 % for G2 or 7 % for G3.

We have been taking the yield curves of various fullerene ions [3]. Geometrical structures and electronic properties of fullerenes have attracted widespread attention because of their novel structures, novel reactivity, and novel catalytic behaviors as typical nanometer-size materials. However, spectroscopic information was very limited in the extreme UV region, owing to difficulties in acquiring enough amount of sample. This situation has been rapidly changed since the start of this century, because the techniques of syntheses, isolation, and purification have been advanced so rapidly that appreciable amount of fullerenes can be readily obtained.



Fig .1 18-m spherical grating monochromator at BL2B.



Fig. 2 End station of BL2B for gas phase spectroscopy of refractory materials.

- [1] M. Ono, H. Yoshida, H. Hattori and K. Mitsuke, Nucl. Instrum. Meth. Phys. Res. A **467-468**, 577 (2001).
[2] H. Yoshida and K. Mitsuke, J. Synchrotron Radiation **5** (1998) 774.
[3] J. Kou, T. Mori, Y. Kubozono and K. Mitsuke, Phys. Chem. Chem. Phys. **7** (2005) 119.

Beamline Specifications

Monochromator	18-m spherical grating Dragon-type
Wavelength Range	6 – 55 nm
Resolution	2000-8000
Experiment	Mass spectrom.; Photoelectron spectrosc.

BL3U

Varied-line-spacing Plane Grating Monochromator for Molecular Soft X-ray Spectroscopy

The beamline BL3U is equipped with an in-vacuum undulator composed of 50 periods of 3.8 cm period length. The emitted photons are monochromatized by the varied-line-spacing plane grating monochromator (VLS-PGM) designed for various spectroscopic investigations in the soft x-ray range including soft x-ray emission studies. Three holographically ruled laminar profile plane gratings are designed to cover the photon energy range from 60 eV to 800 eV. The beamline has two endstations, namely XES setup and Multi-purpose setup. The XES setup is used for soft x-ray emission spectroscopy. The beam is horizontally focused onto the sample position by plane-elliptical mirror, M2X. In the Multi-purpose setup, the beam is focused by the toroidal mirror M2. Between the sample position and M2, the differential pumping is placed.




Fig. 1. Schematic layout (left) and the photography (right) of the BL3U. The distances along the beam from the center of the in-vacuum plane undulator are shown in mm. S1X and M2X can be replaced with the other exit slit S1 so that experiments can be carried out at either the XES or multi-purpose endstation. In the XES setup, the sample is placed at 5-10 mm downstream of S1X.

Beamline Specifications

Monochromator	Varied-line-spacing plane grating monochromator
Energy Range	60-800 eV
Resolution	$E/\Delta E > 10\,000$
Experiment	Soft X-ray spectroscopy (XPS, XES, XAS)
Beam Size (XES Endstation)	Gaussian shape Vertical 5-20 μm ; Horizontal 41 μm (FWHM)

BL4A1

Irradiation Desorption on XeF₂ Gas Etching

BL4A1 has been constructed to perform the synchrotron radiation induced etching of Si and SiO₂ using XeF₂ as an etching gas. This beam-line is composed of a multilayered mirror (MLM) monochromator, a beam condenser system, and a differential pump system. The XeF₂ pressure during the etching will reach to 0.5 Torr, so a differential pump apparatus is installed in the vacuum system and the etching chamber as shown in Fig. 1. The etching chamber is evacuated independently and is designed to achieve high pressure (0.5 Torr) keeping other vacuum system at low pressure ($< 10^{-5}$ Torr) by an aperture flange and a sequence of pressure stages. The condenser mirror focuses the divergent radiation onto the sample surface in the etching chamber, and obtains an extreme higher photon flux can be obtained.




Fig. 1 Section view of differential pump apparatus installed in the etching chamber (reaction cell) and the entrance of beam.



Fig.2 A side view of the end-station at BL4A1.

Beamline Specifications

Monochromator	Multilayered mirror monochromator
Wavelength Range	13.3 ~ 22.5 nm
Resolution	5 ~ 9 eV (FWHM)
Experiment	Irradiation
Miscellaneous	Not-in-use for SR users

BL4A2

SR-induced Photo Etching and CVD Beam Line

This beam line is used for synchrotron radiation (SR)-induced photo-etching and chemical vapor deposition (CVD) experiments. The beam line has no monochromator for high photon flux to irradiate and consists of only two mirrors. One is for focusing and the other is for branching. At the beam line, the gas supply and extinction system is equipped for using legally controlled high pressure gasses such as SiH_4 , Si_2H_6 and GeH_4 . They are commonly used to CVD of semiconductor crystals.

The SR-CVD and photo-etching chambers are connected to the beam line as shown in Fig. 1. In those chambers, infrared reflection absorption spectroscopy (IRRAS) system is installed to study the surface photochemistry on Si surfaces modified with various kinds of molecules.




Figure 1 Top view of BL4A2

Specifications

Spectral range: whole range of synchrotron radiation from UVSOR

BL4B

Varied-line-spacing Plane Grating Monochromator for Molecular Soft X-ray Spectroscopy

The beamline BL4B equipped with a varied-line-spacing plane grating monochromator (VLS-PGM) was constructed for various spectroscopic investigations in a gas phase and/or on solids in the soft X-ray range. Three holographically ruled laminar profile plane gratings with SiO₂ substrates are designed to cover the photon energy range from 25 eV to 800 eV. The gratings with the groove densities of 100, 267, and 800 1/mm cover the spectral ranges of 25-100, 60-300, and 200-1000 eV, respectively, and are interchangeable without breaking the vacuum. Fig. 1 shows the absolute photon flux for each grating measured by a Si photodiode (IRD Inc.), with the entrance- and exit-slit openings set at 50 and 50 μm , respectively. The maximum resolving power ($E/\Delta E$) achieved for each grating is more than 5000.




Fig. 1 Throughput from the VLS-PGM monochromator on BL4B.



Fig. 2 Photo of BL4B.

Beamline Specifications

Monochromator	Varied-line-spacing Plane Grating Monochromator
Energy range	25-1000 eV
Resolution	$E/\Delta E > 5000$ (at maximum)
Experiments	Soft X-ray spectroscopy (mainly, angle-resolved photoion spectroscopy for gaseous targets and photoelectron spectroscopy for gaseous and solid targets)

BL5U

Photoemission Spectroscopy of Solids and Surfaces

This beamline is designed for a high-resolution angle-resolved photoemission study on solids and surfaces with horizontal-linearly and circularly (CW, CCW) polarized synchrotron radiation from a helical undulator. The beamline consists of a Spherical Grating Monochromator with Translational and Rotational Assembly Including a Normal incidence mount (SGM-TRAIN), and a high-resolution angle-resolved photoemission spectrometer.

The SGM-TRAIN is an improved version of a constant-length SGM to aim the following points; (1) covering the wide energy range of 5-250 eV, (2) high energy resolving power, (3) use of linearly and circularly polarized undulator light, (4) reduction of higher order light, and (5) two driving modes (rotation and translation of gratings) by computer control. The second-order light is well suppressed by using laminar profile gratings and combinations of mirrors and gratings.




Fig. 1. Throughput spectra from the SGM-TRAIN monochromator at BL5U.

Beamline Specifications

Monochromator	SGM-TRAIN
Energy Range	5-250 eV
Resolution	$h\nu/\Delta E > 2,000$ for $< 40\mu\text{m}$ slits
Experiment	ARPES, AIPES, XAS
Flux	$< 10^{11}$ photons/s for $< 40\mu\text{m}$ slits (at the sample position)
Main Instruments	Hemispherical photoelectron analyzer (MBS-Toyama 'Peter' A-1), LEED of reverse type (OMICRON), Liq-He flow cryostat (5 – 400 K)




Fig. 2. High-resolution angle-resolved photoemission apparatus at BL5U

BL5B

Calibration Apparatus for Optical Elements and Detectors

BL5B has been constructed to perform calibration measurements for optical elements and detectors. This beamline is composed of a plane grating monochromator (PGM) and three end stations in tandem. The most upstream station is used for calibration measurements of optical elements, the middle one for optical measurements for solids and the last for photo-stimulated desorption experiments. The experimental chamber at the most downstream station is sometimes changed to a chamber for photoemission spectroscopy.

The calibration chamber shown in Fig. 2 is equipped with a goniometer for the characterization of optical elements, which has six-degree-of-freedom; X-Y translation of a sample, and interchange of samples and filters. These are driven by pulse motors in vacuum. Since the polarization of synchrotron radiation is essential for such measurements, the rotation axis can be made in either horizontal or vertical direction (s- or p-polarization).




Fig. 1 Throughput spectra for possible combinations of gratings and mirrors at BL5B measured by a gold mesh.




Fig. 2 A side view of the experimental chamber for calibration measurements.

Beamline Specifications

Monochromator	Plane Grating Monochromator
Energy range	6-600 eV (2-200 nm)
Resolution	$E/\Delta E \sim 500$
Experiments	Calibration of optical elements, absorption of solids, photo-stimulated desorption from rare gas solids

BL6B (IR)

Infrared and Terahertz Spectroscopy of Solids

SR has a good performance (high brilliance and high flux) not only in VUV and SX regions but also in IR and terahertz regions. This beamline covers in the IR and terahertz regions. The previous beamline BL6A1 that has been constructed in 1985 is the pioneer of the infrared SR research. The beamline was upgraded in the spring of 2004 and the name was changed to be BL6B (IR). The front-end part including the bending duct #6 was replaced to a new one with higher acceptance angle using a magic mirror as shown in Fig. 1.

The beamline is equipped with two interferometers, one is Michelson-type (Bruker IFS-66v) and the other Martin-Puplett-type (JASCO FARIS-1), for the wide spectral region from several to 20,000 cm^{-1} ($h\nu =$ several 100 $\mu\text{eV} - 2.5 \text{ eV}$) as shown in Fig. 2. The experimental chamber in which users bring can be equipped at the free port. In the near future, an IR microscope covering down to terahertz region will be set up.




Fig. 1 The design of optics and front end of BL6B.




Fig. 2 Schematic figure of top view of BL6B.

Beamline Specifications

Interferometer	Michelson (Bruker IFS66v), Martin-Puplett (JASCO FARIS-1)
Wavenumber Range (Energy range)	several – 20,000 cm^{-1} , (several 100 $\mu\text{eV} - 2.5 \text{ eV}$)
Resolution in cm^{-1}	0.1 cm^{-1} for IFS66v, 0.25 cm^{-1} for FARIS-1
Experiments	Reflectivity and transmission, Magneto-optics (Microspectroscopy)
Miscellaneous	Users can bring their experimental system in this beamline.

BL7U

Angle-resolved Photoemission of Solids in the VUV region

The beamline 7U is constructed to provide the photon flux with high energy resolution and high flux mainly for high-resolution angle-resolved photoemission spectroscopy of solids. An APPLE-II-type variable polarization undulator is equipped for the light source. The undulator can make high intense VUV light with horizontal/vertical linear and right/left circular polarization. The undulator light is monochromatized by the modified Wadsworth-type monochromator with three gratings ($R = 10$ m; 1200, 2400 and 3600 lines/mm optimized at $h\nu = 10, 20,$ and 33 eV). The energy resolution of light ($h\nu/\Delta h\nu$) is more than 10^4 with the photon flux of more than $10^{11} \sim 10^{12}$ ph/s on samples in the whole energy region.

The beamline has a photoemission end station which equips a 200-mm-radius hemispherical photoelectron analyzer (MB Scientific AB, A-1 analyzer) and a liquid-helium-cooled cryostat with a 6-axes pulse motor control (A-VC Co. Ltd., i-GONIO). The main purpose is to determine the three-dimensional Fermi surface and electronic structure of solids at low temperatures and their temperature dependence to reveal the origin of the physical properties.




Fig.1. Schematic figure of BL7U.

Beamline Specifications

Light source	APPLE-II type undulator ($\lambda_u = 76\text{mm}, N = 36$)
Monochromator	10-m Normal Incidence Monochromator (modified Wadsworth-type)
Photon energy range	6 ~ 40 eV ($\lambda = 30 \sim 200$ nm)
Resolution ($h\nu/\Delta h\nu$)	$1 \times 10^4 \sim 5 \times 10^4$
Photon flux on sample	$\geq 10^{12} \sim 10^{11}$ ph/s (depend on $h\nu$)
Beam size on sample	$200(\text{H}) \times 50(\text{V}) \mu\text{m}^2$
Experiments	Angle-resolved photoemission of solids (MB Scientific A-1 analyzer)

BL8B1

Spherical Grating Monochromator for Soft X-Ray Spectroscopic Studies on Solids and Surfaces

The beamline BL8B1 equipped with a constant-deviation constant-length spherical grating monochromator [1] provides soft X-ray photons in the energy range 30-800 eV with medium energy resolution. The photon energy range is covered by using three gratings (R=15 m; 1080 l/mm, R=15 m; 540 l/mm, and R=7.5m; 360 l/mm) which are interchangeable in vacuum. Figure 1 shows a throughput spectrum measured with the entrance- and exit-slit openings of 10 μm . Under this condition, the achievable resolving power is about 4000 at 400 eV and 3000 at 245 eV, respectively.

An experimental chamber is equipped for conventional measurements of electron yield spectra, or pseudo-photoabsorption spectra, under a $\sim 1 \times 10^{-6}$ Torr vacuum condition.

[1] Hiraya et al., Rev. Sci. Instrum. **66** (1995) 2104.



Fig. 1 Photo of BL8B1.




Fig. 2 Throughput of the monochromator at BL8B1.

Beamline Specifications

Monochromator	Constant-deviation constant-length spherical grating type
Energy range	30-800 eV
Resolution	$E/\Delta E = 4000$ at 400 eV and 3000 at 245 eV
Experiments	Photoabsorption spectroscopy, electron spectroscopy and electron-ion coincidence spectroscopy for solids and surfaces

BL8B2

Angle-Resolved Ultraviolet Photoelectron Spectrometer for Solids

BL8B2 is a beamline for angle-resolved ultraviolet photoemission spectroscopy (ARUPS) system which is designed for measuring various organic solids such as molecular crystals, organic semiconductors, and conducting polymers. This beamline consists of a plane-grating monochromator (PGM), a sample preparation chamber with a fast entry Load-Lock chamber, a measurement chamber with an accurate for temperature dependence (base pressure 1×10^{-10} Torr), a cleaning chamber (base pressure 1×10^{-10} Torr), and a sample evaporation chamber (base pressure 3×10^{-10} Torr). The cleaning chamber is equipped with a back-view LEED/AUGER, an ion gun for Ar^+ sputtering, and an infrared heating unit. The PGM consists of premirrors, a plane grating, focusing mirror, and a post-mirror, with an exit slit. It covers the wide range from 2 to 130eV with exchanging two gratings (G1: 1200l/mm, G2: 450l/mm) and five cylindrical mirrors. The toroidal mirror focuses the divergent radiation onto the sample in the measurement chamber. The spot size of the zeroth-order visible light at the sample surface is about $1 \times 1 \text{mm}^2$. Figure 1 shows the throughput spectra of PGM (slit=100 μm). The energy resolution at a slit width of 100 μm was found to be $E/\Delta E = 1000$ in the wavelength range from 2 to 130eV. A hemi-spherical electron energy analyzer of 75mm mean radius with an angular resolution less than 2° can be rotated around vertical and horizontal axes. The sample mounted on a manipulator can be also rotated around two axes.




Fig. 1 Throughput spectra of plane-grating monochromator at BL8B2 (slit=100 μm).



Fig. 2 A photo of BL8B2.

Beamline Specifications

Monochromator	Plane-grating monochromator
Wavelength Range	9- 600nm
Resolution	$E/\Delta E = 1000$
Experiment	Angle-resolved ultraviolet photoemission spectroscopy